Notification of a Body in the framework of a technical harmonization directive

From: Lægemiddelstyrelsen / Danish To: European Commission

Medicines Agency Axel Heides Gade 1 DK-2300 Københvn S

Denmark

GROWTH Directorate-General 200 Rue de la Loi,

B-1049 Brussels.

Other Member States

Reference: Legislation: Regulation (EU) 2017/745 on medical devices

Body name, address, telephone, fax, email, website:

TÜV SÜD Danmark Strandvejen 125 2900 Hellerup Denmark +45 4477 0010

info.dk@tuvsud.com

https://www.tuvsud.com/en-gb/country/denmark

Body info: NB 2443

Body:

Created: 2024-04-05 **Last update:** 2024-04-05

Tasks performed by the Body:

Last approval date: 2024-04-06

Products	Procedures	Articles /Annexes	Conditions
DESIGN AND INTENDED PURPOSE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -2. Active non-implantable devices for imaging, monitoring and/or diagnosis -MDA 0202 Active non- implantable imaging devices utilising non-ionizing radiation	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Limited to ultrasound technology
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -2. Active non-implantable devices for imaging, monitoring and/or diagnosis -MDA 0203 Active non- implantable devices for monitoring of vital physiological parameters	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	

Products	Procedures	Articles /Annexes	Conditions
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -A. Active devices -3. Active non-implantable therapeutic devices and general active non-implantable devices -MDA 0305 Active non- implantable devices for stimulation or inhibition	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Limited to infusion systems
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Limited to hearing aids

Products	Procedures	Articles /Annexes	Conditions
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Excluding surgical robotic
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Excluding active limb prosthesis
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
DESIGN AND INTENDED PURPOSE OF THE DEVICE	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	Excluding bone substitutes and bone cements

Products	Procedures	Articles /Annexes	Conditions
Floducis	Procedures	Articles /Aririexes	Conditions
L COREO REEL FOTING THE	On the second se	A	
DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1201 Non-active non- implantable devices for anaesthesia, emergency and intensive care	Conformity assessment based on assessment of technical documentation	Annex IX(II)	
-I. CODES REFLECTING THE DESIGN AND INTENDED PURPOSE OF THE DEVICE -B. Non-active devices -2. Non-active non-implantable devices -MDN 1203 Non-active non- implantable guide catheters, balloon catheters, guidewires, introducers, filters, and related tools	Conformity assessment based on a quality management system Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	
-2. Non-active non-implantable devices -MDN 1208 Non-active non-implantable instruments	Conformity assessment based on assessment of technical documentation	Annex IX(II)	
	Conformity assessment based on assessment of technical documentation	Annex IX(I) Annex IX(II)	

H	Horizontal technical competences	Limitations
MDS 1001 substances:	Devices incorporating medicinal	
MDS 1002 cells of human	Devices manufactured utilising tissues or origin, or their derivatives:	
MDS 1003 cells of animal	Devices manufactured utilising tissues or origin, or their derivatives:	
MDS 1004 Devices which are also machinery as defined in point (a) of the second paragraph of Article 2 of Directive 2006/42/EC of the European Parliament and of the Council (1):		

Н	orizontal technical competences	Limitations
MDS 1005	Devices in sterile condition:	Including aseptic processing, ethylene oxide gas sterilisation (EOG), low temperature steam and formaldehyde sterilisation, moist heat sterilisation, radiation sterilisation (gamma, x-ray, electron beam), sterilisation with hydrogen peroxide, sterilisation with liquid chemical sterilising agents, thermic sterilisation with dry heat
MDS 1008	Devices utilising biologically active	
coatings and/or	r materials or being wholly or mainly	
	cally dispersed in the human body or are	
	dergo a chemical change in the body:	
MDS 1009	Devices incorporating software/utilising	
software/contro	olled by software, including devices	
intended for co	ntrolling, monitoring or directly influencing	
the performance	e of active or active implantable devices:	
MDS 1010	Devices with a measuring function:	
MDS 1011	Devices in systems or procedure packs:	
MDS 1013 devices:	Class III custom-made implantable	
MDT 2001 processing:	Devices manufactured using metal	
MDT 2002 processing:	Devices manufactured using plastic	
MDT 2007 regarding the p	Devices which require knowledge roduction of pharmaceuticals:	
MDT 2008 and associated	Devices manufactured in clean rooms controlled environments:	
MDT 2009 of materials of	Devices manufactured using processing human, animal, or microbial origin:	
MDT 2010 components in	Devices manufactured using electronic cluding communication devices:	
MDT 2011 including labell	Devices which require packaging, ing:	
MDT 2012 refurbishment:	Devices which require installation,	